

Technical Data Sheet


Article-No.: **890012**




Article-Description: **Nitrile examination glove, powderfree, blue, non sterile**

PRODUCTDESCRIPTION					
material	<input type="checkbox"/> Latex	<input checked="" type="checkbox"/> Nitrile	<input type="checkbox"/> Vinyl	<input type="checkbox"/> Polyethylen (PE)	<input type="checkbox"/> cotton
colour	<input type="checkbox"/> white	<input checked="" type="checkbox"/> blue	<input type="checkbox"/> mint	<input type="checkbox"/> black	<input type="checkbox"/> light pink
characteristic	<input type="checkbox"/> pre-powdered	<input checked="" type="checkbox"/> powder free	<input checked="" type="checkbox"/> non sterile	<input checked="" type="checkbox"/> ambidextrous	
surface	<input type="checkbox"/> textured	<input checked="" type="checkbox"/> textured fingertips	<input type="checkbox"/> embossed	<input checked="" type="checkbox"/> double polymer coated	

SIZES						
	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)	XXL (10-11)
width	< 80 mm	80 mm ± 10 mm	95 mm ± 10 mm	110 mm ± 10 mm	115 mm ± 10 mm	125 mm ± 5 mm
length	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm

THICKNESS		
cuff	palm	fingertips
0,06 mm	0,08 mm	0,09 mm

REGULATORY AFFAIRS							
PPE-Regulation (EU) 2016/425	<input type="checkbox"/> Category I	<input type="checkbox"/> Category II	<input checked="" type="checkbox"/> Category III	<input type="checkbox"/> no PPE-article			
MD-Regulation (EU) 2017/745	<input checked="" type="checkbox"/> Class 1	<input type="checkbox"/> Class 2a	<input type="checkbox"/> Class 2b	<input type="checkbox"/> Class 3	<input type="checkbox"/> sterile	<input type="checkbox"/> with measuring function	<input type="checkbox"/> no medical device
This product is in compliance with the rules for protective gloves, which are intended to come into contact with food: Regulation (EC) 1935/2004					<input checked="" type="checkbox"/> yes	<input type="checkbox"/> no	

STANDARDISATION							
EN 388:2016 mechan. risks	abrasion resistance	blade cut resistance Coupe-Test	tear resistance	puncture resistance	blade cut resistance TDM-Test	impact test	
Level	not applicable						
EN 374-1:2016 chem. Risks	chemical		code letter	level	breakthrough time	degradation	ISO 374-1 / Typ B 
EN 374-4:2013 Degradation	Sodium hydroxide 40%		K	6	> 480 min	-11,5 %	
	Hydrogen Peroxide 30%		P	4	> 120 min	19,1 %	
	Formaldehyde 37%		T	6	> 480 min	27,2 %	
EN 374-5:2016 microorganism tightness	The glove is tight against microorganisms (viral, bacteria and fungi). Test according to ISO 16604 - method B						
EN 420:2010 protective gloves	The glove meets the requirements according to EN 420:2010						
EN 455 medical gloves for single use	The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4						EN 455
EN 455-1:2000 freedom from holes	The glove has an AQL <1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1, general Inspection Level I)						AQL 1.5

ADDITIONAL TESTED CHEMICALS (contact us, if you need more breakthrough-times)							
cytostatic drug	Cisplatin 1,0 mg/ml	Cyclophosphamide 20 mg/ml	Darcarbazine	Etoposide 20 mg/ml	Melphalan	Methotrexate 25 mg/ml	Ifosamide
Level	5	5	5	5	5	5	5